

VNA parameter	Raman DTS	FBG sensors
Start frequency/frequency step	30 kHz	500 kHz
Stop frequency	99.99 MHz	100 MHz
IF bandwidth	30 Hz	10 kHz
Number of frequency samples	3333	200
Averaging	10	–
Total measurement time	≈ 20 min	≈ 2 min